STABILITY: Stable

2 to 10 field blanks

BLANKS:

EXCITATION λ: 380 nm

Detection \lambda range; λ_{max} : 400 – 700 nm; 475 nm

ACCURACY

CALIBRATION: Elemental beryllium in detection solution

RANGE STUDIED: 0.02 to 12.0 µg/wipe

RANGE: 0.06 to 6 μ g/wipe

BIAS: none identified

ESTIMATED LOD:0.02 μg/wipe

OVERALL

PRECISION (,): $0.021 @ \sim 0.2 \mu g/wipe (n = 15);$

w/ C 13 1

 $\sim 3\mu g/filter (n = 6)$

0.076 @ ~1.5 µg/wipe (n = 6); 0.052 @

PRECISION ($\hat{\mathbf{S}}_{rT}$): 0.057

ACCURACY: 11.6 (6.4 – 16.8, 95% C.I.)

APPLICABILITY: The working range of the method is 0.02 to $6.0 \,\mu\text{g}/100 \,\text{cm}^2$ for surface wipe samples. The analysis is for total beryllium and is not compound specific.

INTERFERENCES: Minor interference (<10%) from Fe can result if iron concentrations are high. Samples high in iron demonstrate a yellow or gold coloration. This interference can be minimized by allowing the solution to sit for at least four hours, during which time the solution clears and then filtering the sample extract before use,

OTHER METHODS: Method 7300 (hot plate or microwave digestion and inductively coupled plasma atomic emission spectrometry) is an alternative (reference) procedure for the determination of elemental beryllium. ASTM method D7202-05 is a similar procedure to detect elemental beryllium by fluorescence.

REAGENTS:

- Dissolution solution: 1% Ammonium bifluoride
 (prepared from dissolving 1g ammonium bifluoride in 100ml of deionized water).
- Detection solution: 63.4 μM hydroxyl benzoquinoline sulfonate (HBQS) [1] / 2.5 mM ethylene diamine tetraacetic acid (EDTA) / 50.8 mM lysine monohydrochloride (pH adjusted to 12.85 with 10 M NaOH).
- 3. Water, deionized.
- 4. Beryllium spectroscopic standard solution,~1,000 μg/ml (commercially available).
- Spiked beryllium media (commercially available)

* See SPECIAL PRECAUTIONS

EQUIPMENT:

- Wipe material, cellulosic, 47-mm minimum diameter, wetted
 - NOTE: Polyvinyl alcohol (PVA) media are unsuitable for this method.
- 2. Templates, disposable reusable, 100 cm² minimum area
- 3. Tape, masking
- 4. Portable ultraviolet/visible (UV/Vis)
 fluorometer, with irradiance excitation lamp (λ
 = 380 nm) and time-integrating visible detector
 (400 700 nm, λ_{max} ≈ 475 nm)
- 5. Mechanical agitator, shaker or rotator
- Fluorescence cuvettes, disposable, 10-mm diameter, transparent to UV/Vis radiation
- 7. Centrifuge tubes, plastic, 15-ml
- Syringe filters, 0.45-µm nylon, 13- or 25-mm diameter, in plastic housings.
 - NOTE: Polytetrafluoroethylene (PTFE) filters are unsuitable for this method.
- Pipettors, mechanical, of assorted sizes as needed
- Pipet tips, plastic, disposable, of assorted sizes as needed
- 11. Labware, plastic (e.g., beakers, flasks,

graduated cylinders, etc.), of assorted sizes as needed

- 12. Tweezers, plastic or plastic-coated
- 13. Laboratory wipes
- Personal protective wear, e.g., respirators, masks, gloves, lab coats, safety eyewear, etc. as needed

special precautions: Wear appropriate personal protection during sampling activities and analysis. Perform sample preparation and analysis in a clean well ventilated area that is well removed from any possible beryllium contamination. It is ESSENTIAL that suitable personal protective equipment, including suitable gloves, eye protection, laboratory coat, etc. is used when working with the chemicals. Any area affected by the detection or dissolution solution must be immediately washed with plenty of water. Ammonium bifluoride will etch glass, so it is essential that all NH₄HF₂ solutions are contained in plastic labware. Avoid exposure by contact with the skin or eyes, or by inhalation of the vapor.

8

9

10

11

SAMPLING:

- 1. Don a clean pair of gloves.
- 2. Demarcate the sampling area (100 cm² minimum) using a clean template or tape. If a template is used, tape the outside edges of the template to the surface to prevent its moving during sampling.

14	3.	Using a clean wipe, sample the demarcated area using the procedure described in ASTM D6966
15		[2].
16	4.	Place each wipe sample into labeled 15-ml plastic centrifuge tubes.
17		
18	SAMP	LE PREPARATION:
19	5.	Add 5 ml of the dissolution solution (1% ammonium bifluoride) to each 15-ml centrifuge tube
20		containing surface swipe sample, and cap each tube.
21	6.	Place each tube into a mechanical rotator, and rotate for at least 30 min.
22		NOTE: Rotator may also be substituted by a shaker or an agitator as long as the dissolution
23		solution wets the wipe well. Sonication has also been shown to be effective. Dissolution of
24		refractory material such as beryllium oxide by heating the solution to 80°C for 30 minutes
25		without agitation has been shown to be effective.
26	7.	Into clean beakers, flasks, or tubes, filter each solution with a nylon syringe filter.
27	8.	Into cuvettes containing 1.9 ml of detection solution, pipet 0.1 ml of each sample filtrate. Cap and
28		mix briefly.
29		NOTE: If high iron or titanium concentration is suspected or is evident (owing to the appearance
30		of suspended precipitate), allow the solution to settle, or filter the solution using a nylon
31		syringe filter.
32		NOTE: The stability of the detection and the dissolution solution is more than one year and of the
33		mixed measurement solution comprising both is greater than 30 days. The solutions must
34		be kept in sealed containers and the detection and mixed solutions must be stored away
35		from light.
36		NOTE: Alternative ratios of dissolution solution comprising beryllium and detection solution may
37		be used for analyzing alternative range of beryllium concentration.
38		
39	CALIB	RATION AND QUALITY CONTROL:
40	9.	Calibrate the fluorometer according to the manufacturer's recommendations.
41		NOTE: Beryllium stock standard solutions are made up using Beryllium spectrometric standards
42		diluted into 1% ammonium bifluoride. Calibration check standards are then prepared by

43	adding 0.1 ml of beryllium stock standards into 1.9 ml of detection solution (20-fold
44	dilution). A recommended series of standard stock solutions are 800 ppb, 200 ppb, 40
45	ppb, 10 ppb and 0 ppb to measure a range of 0.2 μg to 4 μg of beryllium on the sampling
46	media.
47	NOTE: If alternative ratios of dissolution solution comprising beryllium and detection solution are
48	used for sample preparation, then a similar ratio for calibration is required.
49	·
50	10. Analyze a calibration standard, a reagent blank, and a media blank at least once every 20
51	samples. Ensure that the concentration range of the calibration standards spans the measured
52	beryllium levels in the samples.
53	11. Check recoveries with at least two media spikes per ten samples.
54	NOTE: It is recommended to use beryllium oxide (BeO) for media spikes.
55	
56	MEASUREMENT:
57	
58	12. For each sample, obtain the fluorescence spectrum using a 380-nm excitation lamp and visible
59	(400-700 nm) detector (follow instrument manufacturer's instructions).
60	13. If the fluorescence values for any of the samples are above the range of the calibration
61	standards, dilute the solutions with detection solution, reanalyze, and apply the appropriate
62	dilution factor in subsequent calculations.
63	
64	CALCULATIONS:
65	14. Obtain the solution concentration for each sample, C_s ($\mu g/I$), and the average media blank, C_b
66	(μg/l).
67	15. Using the solution volumes of sample, V_s (ml), and media blank, V_b (ml), calculate the
68	concentration, C ($\mu g/100 \text{ cm}^2$) of Be in the sample of area A, while accounting for the dilution
69	factor DF. The calculations below give the amount of beryllium in the wipe. It is assumed that 100
70	cm ² area was wiped to collect the sample. If the area wiped is different from this number then the
71	calculations need to be adjusted so that the results are normalized to 100 cm ² area

$$C = DF \times \frac{\left[C_{s} V_{s} - C_{b} V_{b}\right]}{A}, \mu g/100 \text{cm}^{2}$$

NOTE: The table below can be used for correlating the amount of beryllium in the solution to the beryllium in the sampling media.

Preparation of Standard	Final concentration of beryllium	Corresponding amount of
Solutions	(ppb) in calibration standard	beryllium in the media*
	solutions	
0.1 ml of 0 ppb standard +	0.0	Corresponds to 0.00 μg Be on
1.9 ml of detection solution		media
0.1 ml of 10 ppb standard +	0.5	Corresponds to 0.05 μg Be on
1.9 ml of detection solution		media
0.1 ml of 40 ppb standard +	2.0	Corresponds to 0.2 μg Be on
1.9 ml of detection solution		media
0.1 ml of 200 ppb standard +	10.0	Corresponds to 1 µg Be on
1.9 ml of detection solution		media
0.1 ml of 800 ppb standard +	40.0	Corresponds to 4 μg Be on
1.9 ml of detection solution		media

^{*}Incorporating sample dilution factor for 5 ml of dissolution solution; note that volumes other than 5 ml will require a different appropriate dilution factor.

EVALUATION OF METHOD:

The method was evaluated in accordance with published guidelines [3]. Experiments were conducted using an Ocean Optics® portable fluorescence device with the following components:

USB 200 spectrometer with spectral grating #2 (UV/Vis 600)

86 LS-1 lamp (380-nm) in LS-450 housing

87	UV-2 casting
88	OFLV linear filter 200-850
89	L2 collection lens and slit-200
90	
91	Tests were carried out in relative irradiance mode using 2- or 5-sec integration times.
92	
93	The method was evaluated using beryllium oxide spiked onto Whatman #541 cellulose and nylon filters at
94	levels of 0, 0.02, 0.1, 0.2, 0.3, 0.4, 1.5, 3.0, and 6.0 μg (five samples at each level).
95	
96	The method was also field-tested using real-world samples (collected using Whatman #541 cellulose and
97	nylon filters) obtained at U.S. Department of Energy sites suspected to be contaminated with beryllium.
98	Field samples measured by portable fluorometry were also analyzed using hot plate digestion and ICP-
99	AES analysis; the latter served as a reference analytical method. Sample loadings observed ranged from
100	below detection limit (<0.02 μg/sample) to ~12 μg/sample.
101	

L02	Long-term stability of samples was verified from spikes (n = 30) of 0.1 µg Be on Whatman #541 cellulose
L03	and nylon filters. Samples were analyzed at day one (n = 12) and then one week (n = 6), ten days (n = 3),
L04	two weeks (n = 3), three weeks (n = 3), and one month (n = 3) after spiking. No diminution of fluorescence
L05	signal was observed from samples prepared and analyzed after having been stored for up to thirty days.
L06	
L07	Interference tests were carried out using solutions of 0 nM, 100 nM, and 1.0 μ M Be in the presence of 0.4
L08	mM Al, Ca, Co, Cu, Fe, Ti, Li, Ni, Pb, Sn, U, V, W, and Zn (separate experiments were carried out for each
L09	potential interferant).
L10	
111	REFERENCES:
L12	[1] Matsumiya H, Hoshino H, Yotsuyanagi T (2001), A novel fluorescence reagent, 10-
L13	hydroxybenzo(h)quinoline-7-sulfonate, for selective determination of beryllium(II) ion at pg cm ⁻¹ levels.
L14	Analyst 126: 2082-2086.
L15	[2] ASTM D6966 (2003), Standard Practice for Collection of Surface Wipe Samples for Subsequent
L16	Determination of Metals. American Society for Testing and Materials (ASTM): West Conshohocken,
L17	PA.
L18	[3] Kennedy ER, Fischbach TJ, Song R, Eller PM, Shulman SA (1995), Guidelines for Air Sampling and
L19	Analytical Method Development and Evaluation. CDC/NIOSH: Cincinnati, OH; DHHS (NIOSH) Publ.
L20	No. 95-117.
L21	[4] Minogue EM, Ehler DS, Burrell AK, McCleskey TM, Taylor TP [2005]: Development of a new
L22	fluorescence method for the detection of beryllium on surfaces. J. ASTM Int. 2(9), 10 pp.; Paper ID
L23	JAI13168
L24	[5] Ashley K, McCleskey TM, Brisson MJ, Goodyear G, Cronin J, Agrawal A [2005]: Interlaboratory
L25	evaluation of a portable fluorescence method for the measurement of trace beryllium in the workplace.
L26	J. ASTM Int. 2 (9), 8 pp.; Paper ID JAI13156.
L27	
L28	METHOD DEVELOPED BY:
129	T. Mark McCleskey, Los Alamos National Laboratory, Los Alamos, NM

130

131 **METHOD WRITTEN BY:**

- 132 T. Mark McCleskey, Los Alamos National Laboratory, Los Alamos, NM
- Kevin Ashley, Ph.D., CDC/NIOSH, Division of Applied Research and Technology
- 134